

Ten-year Outcomes of the PARTNER 2 Intermediate-Risk Studies:

A Propensity-matched Analysis of P2S3i TAVR and P2A Surgery

Raj Makkar, MD

on behalf of The PARTNER 2 Trial Investigators

TCT | San Francisco | October 27th, 2025



Disclosure Statement of Financial Interest

Raj Makkar, MD

Within the past 12 months, I or my spouse/partner have had a financial interest/arrangement or affiliation with the organization(s) listed below.

Affiliation/Financial Relationship

- Grant/Research Support
- Consultant fee/Honoraria

Company

- Edwards Lifesciences, St Jude Medical
- Edwards Lifesciences, Abbott Vascular, Cordis Corporation, Medtronic

Background (1)




- The third-generation SAPIEN 3 transcatheter heart valve was introduced shortly after PARTNER 2A (P2A) trial enrollment.
- The PARTNER 2 SAPIEN 3 Intermediate-risk Registry (P2S3i) was subsequently designed using **the same eligibility criteria** as the P2A trial.
- P2S3i patients were **prospectively enrolled** to undergo S3 TAVR for a **pre-specified comparison** to the surgical arm of the P2A trial.


Background (2)

Valve
Technology


SAPIEN



SAPIEN XT



SAPIEN 3



Sheath
Compatibility

22-24F



16-20F



14-16F



Available
Valve Sizes



23 mm



26 mm



23 mm



26 mm



29 mm



20 mm



23 mm



26 mm



29 mm

Background (3)



1-year Outcomes

THE LANCET
Transcatheter aortic valve replacement versus surgical valve replacement in intermediate-risk patients: a propensity score analysis

Vinod H Thourani, Susheel Kodali, Raj R Makkar, Howard C Herrmann, Mathew Williams, Vasilis Babaliaros, Richard Smalling, Scott Lim, S Chris Malaisrie, Samir Kapadia, Wilson Y Szeto, Kevin L Greason, Dean Kereiakes, Gorav Ailawadi, Brian K Whisenant, Chandan Devireddy, Jonathon Leipsic, Rebecca T Hahn, Philippe Pibarot, Neil J Weissman, Wael A Jaber, David J Cohen, Rakesh Suri, E Murat Tuzcu, Lars G Svensson, John G Webb, Jeffrey W Moses, Michael J Mack, D Craig Miller, Craig R Smith, Maria C Alu, Rupa Parvataneni, Ralph B D'Agostino Jr, Martin B Leon

Thourani VH et al. Lancet 2016; 387:2218-2225.

5-year Outcomes

Outcomes of SAPIEN 3 Transcatheter Aortic Valve Replacement Compared With Surgical Valve Replacement in Intermediate-Risk Patients 

Mahesh V. Madhavan, MD, MS,^{a,b} Susheel K. Kodali, MD,^a Vinod H. Thourani, MD,^c Raj Makkar, MD,^d Michael J. Mack, MD,^e Samir Kapadia, MD,^f John G. Webb, MD,^g David J. Cohen, MD, MSc,^{b,h} Howard C. Herrmann, MD,ⁱ Mathew Williams, MD,^j Kevin Greason, MD,^k Philippe Pibarot, DVM, PhD,^l Rebecca T. Hahn, MD,^{a,b} Wael Jaber, MD,^f Ke Xu, PhD,^m Maria Alu, MS,^{a,b} Craig R. Smith, MD,^a Martin B. Leon, MD^{a,b}

Madhavan MV et al. JACC 2023; 82:109-123.

Here, we present 10-year outcomes of intermediate-risk patients with symptomatic, severe AS who underwent SAPIEN 3 TAVR (P2S3i) or surgery (P2A) using a propensity-matched analysis.


The PARTNER 2A and S3i Trials

Study Designs



Symptomatic Severe Aortic Stenosis

Intermediate Risk Assessment by Heart Valve Team (STS score 4 - 8%)



PARTNER 2 – 10-year Follow-up



Trial Enrollment:

- P2A: 2011 – 2013
- P2S3i: 2014

Reconsent required for follow-up extension to 10 years per FDA request



A Vital Status Sweep (VSS) was performed by sites using patient/family phone calls, publicly-available data, and/or medical records in patients who withdrew, were lost to follow-up, did not reconsent, or missed a visit

Study Methodology



- P2S3i TAVR patients were propensity-score matched 1:1 to P2A surgical patients using the **same methodology as the 5-year analysis.**¹
- Matching was performed in the **valve-implant populations** of both studies.
- **Key endpoints available include:**
 - All-cause death (with VSS)
 - AV reintervention
 - Mean gradient
 - PVR

Assessed by the
**same echo core
lab** for both studies

Patient Disposition

Available for All-cause Mortality Analysis



THE
PARTNER II
TRIAL

Study exits before
treatment (N=9)

Enrolled in P2S3i Registry: N = 1078

Valve Implant Population: N = 1069

TF TAVR
N = 943

TA/TAo TAVR
N = 126

5-year follow-up
N = 897 (83.9%)

Study exits before
10 years (N=422)

Did not consent
(n=207)

Total Study Exits
(N=431)

10-year follow-up
N = 647 (60.5%)

10-year follow-up with VSS
N = 959 (89.7%)

Randomized to Surgery in P2A: N = 1021

Study exits before
treatment (N=85)

Valve Implant Population: N = 936

Suitable for TF
N = 719

Suitable for TA/TAo
N = 217

5-year follow-up
N = 817 (87.3%)

Did not consent
(n=270)

Study exits before
10 years (N=430)

10-year follow-up
N = 506 (54.1%)

Total Study Exits
(N=515)

10-year follow-up with VSS
N = 838 (89.5%)

Baseline Characteristics


Unmatched



Characteristic	TAVR (n = 1069)	Surgery (n = 936)	P-value
Age, y	81.9 ± 6.6	81.6 ± 6.7	0.35
Male	61.6	54.7	0.001
Annulus diameter, mm	21.9 ± 2.2	21.5 ± 2.1	<0.0001
STS Score, %	5.3 ± 1.3	5.8 ± 1.9	<0.0001
NYHA Class III/IV	72.7	75.9	0.06
Hypertension	92.1	94.8	0.01
CAD	69.7	66.4	0.12
Prior CABG	28.1	25.5	0.22
Prior atrial fibrillation	35.7	34.8	0.66
PVD	28.3	31.9	0.02
Renal insufficiency	7.6	5.5	0.02
LVEF, %	58.5 ± 13.4	55.4 ± 11.8	<0.0001
≥ Moderate MR	8.8	18.2	<0.0001

All-cause Mortality

Unmatched




No. at Risk											
S3 TAVR	1069	987	905	811	698	598	482	379	279	201	107
Surgery	936	813	763	701	624	503	396	327	254	184	120

Baseline Characteristics After Matching



Characteristic	TAVR (n = 783)	Surgery (n = 783)	P-value
Age, y	81.7 ± 6.7	81.5 ± 6.8	0.60
Male	58.0	57.0	0.80
Annulus diameter, mm	21.7 ± 2.2	21.7 ± 2.1	0.69
STS Score, %	5.5 ± 1.3	5.5 ± 1.5	0.74
NYHA Class III/IV	74.5	74.6	0.95
Hypertension	93.5	93.9	0.76
CAD	68.6	67.6	0.66
Prior CABG	27.1	25.9	0.61
Prior atrial fibrillation	35.1	34.0	0.63
PVD	29.9	29.5	0.87
Renal insufficiency	6.0	6.0	>0.99
LVEF, %	57.3 ± 14.0	57.0 ± 10.7	0.66
≥ Moderate MR	11.6	12.5	0.62

All-cause Mortality After Matching




No. at Risk											
S3 TAVR	783	726	671	605	515	443	352	276	206	151	80
Surgery	783	682	645	591	523	421	335	281	222	162	103

All-cause Mortality by Anatomical Access Route




TF Access
N = 1,662 (82.9%)



No. at Risk

S3 TAVR	656	607	557	501	426	362	292	233	174	120	61
Surgery	656	578	544	496	437	355	289	244	190	138	89


TA/TAo Access
N = 343 (17.1%)



No. at Risk

S3 TAVR	113	95	83	75	68	63	45	34	23	18	11
Surgery	113	98	92	86	76	58	43	36	28	21	14

AV Reintervention



No. at Risk


S3 TAVR	783	712	647	566	458	310	165	125	101	75	32
Surgery	783	667	613	544	472	296	134	111	90	68	29

Types and Reasons for AV Reintervention



Outcome, No. of Events at 10 years	0 – 5 Years		> 5 – 10 Years	
	S3 TAVR (N=783)	Surgery (N=783)	S3 TAVR (N=165)	Surgery (N=134)
Total Aortic Valve Reintervention	10 pts (11 events)	5 pts (5 events)	2 pts (2 events)	3 pts (3 events)
Reintervention Reason				
Restenosis	1	0	1	2
Aortic Regurgitation	10	1	1	1
Endocarditis	0	4	0	0
Reintervention Type				
Valve-in-valve	8	0	1	3
Surgical Explant	2	5	1	0
BAV	1	0	0	0

Echocardiography in Survivors: Mean Gradient



No. of Echos

S3 TAVR

769

645

552

450

348

269

87

32

Surgery

767

535

458

384

325


249

68

30


Core-lab adjudicated; Patients with explants/VIVs were censored after reintervention

Echocardiography in Survivors: Total Aortic Regurgitation



Core-lab adjudicated; Patients with explants/VIVs were censored after reintervention

Echocardiography in Survivors: Paravalvular Regurgitation



Core-lab adjudicated; Patients with explants/VIVs were censored after reintervention

Study Limitations



- This study compared SAPIEN 3 TAVR to surgery using a propensity-matched analysis; as it was **not a randomized** trial, unmeasured confounders may influence results.
- There was **significant missing data** at 10 years due to the requirement for patient **reconsent** for extended follow-up, disproportionate reconsent and study withdrawal, **loss to follow-up**, and the **competing risk of death** in this elderly population.

Conclusion



At 10 years of follow-up in intermediate-risk patients with symptomatic, severe aortic stenosis, we observed similar mortality and valve durability in SAPIEN 3 TAVR and surgery.

Thank you!

Important Safety Information

Edwards SAPIEN 3, Edwards SAPIEN 3 Ultra, and Edwards SAPIEN 3 Ultra RESILIA Transcatheter Heart Valve System

Indications: The Edwards SAPIEN 3, SAPIEN 3 Ultra, and SAPIEN 3 Ultra RESILIA Transcatheter Heart Valve system is indicated to reduce the risks associated with progression from asymptomatic to symptomatic severe native calcific aortic stenosis in patients who are judged by a heart team to be appropriate for transcatheter heart valve replacement therapy.

The Edwards SAPIEN 3, SAPIEN 3 Ultra, and SAPIEN 3 Ultra RESILIA Transcatheter Heart Valve system is indicated for relief of aortic stenosis in patients with symptomatic heart disease due to severe native calcific aortic stenosis who are judged by a Heart Team, including a cardiac surgeon, to be appropriate for the transcatheter heart valve replacement therapy.

The Edwards SAPIEN 3, SAPIEN 3 Ultra, and SAPIEN 3 Ultra RESILIA Transcatheter Heart Valve system is indicated for patients with symptomatic heart disease due to a failing (stenosed, insufficient, or combined) surgical or transcatheter bioprosthetic aortic valve, or a native mitral valve with an annuloplasty ring who are judged by a heart team, including a cardiac surgeon, to be at high or greater risk for open surgical therapy (i.e., predicted risk of surgical mortality $\geq 8\%$ at 30 days, based on the Society of Thoracic Surgeons (STS) risk score and other clinical co-morbidities unmeasured by the STS risk calculator).

The Edwards SAPIEN 3, SAPIEN 3 Ultra, and SAPIEN 3 Ultra RESILIA Transcatheter Heart Valve system is indicated for patients with symptomatic heart disease due to a failing (stenosed, insufficient, or combined) surgical bioprosthetic mitral valve who are judged by a heart team, including a cardiac surgeon, to be at intermediate or greater risk for open surgical therapy (i.e., predicted risk of surgical mortality $\geq 4\%$ at 30 days, based on the Society of Thoracic Surgeons (STS) risk score and other clinical co-morbidities unmeasured by the STS risk calculator).

Contraindications: The valves and delivery systems are contraindicated in patients who cannot tolerate an anticoagulation/antiplatelet regimen or who have active bacterial endocarditis or other active infections, or who have significant annuloplasty ring dehiscence.

Warnings: Observation of the pacing lead throughout the procedure is essential to avoid the potential risk of pacing lead perforation. There may be an increased risk of stroke in transcatheter aortic valve replacement procedures, as compared to balloon aortic valvuloplasty or other standard treatments in high or greater risk patients. The devices are designed, intended, and distributed for single use only. **Do not resterilize or reuse the devices.** There are no data to support the sterility, nonpyrogenicity, and functionality of the devices after reprocessing. Incorrect sizing of the valve may lead to paravalvular leak, migration, embolization, residual gradient (patient-prosthesis mismatch), and/or annular rupture. Accelerated deterioration of the valve due to calcific degeneration may occur in children, adolescents, or young adults and in patients with an altered calcium metabolism. Prior to delivery, the valve must remain hydrated at all times and cannot be exposed to solutions other than its shipping storage solution and sterile physiologic rinsing solution. Valve leaflets mishandled or damaged during any part of the procedure will require replacement of the valve. Caution should be exercised in implanting a valve in patients with clinically significant coronary artery disease. Patients with pre-existing prostheses should be carefully assessed prior to implantation of the valve to ensure proper valve positioning and deployment. Do not use the valve if the tamper-evident seal is broken or the storage solution does not completely cover the valve (SAPIEN 3 and SAPIEN 3 Ultra only), the temperature indicator has been activated, the valve is damaged, or the expiration date has elapsed. Do not mishandle the delivery system or use it if the packaging or any components are not sterile, have been opened or are damaged (e.g., kinked or stretched), or if the expiration date has elapsed. Use of excessive contrast media may lead to renal failure. Measure the patient's creatinine level prior to the procedure. Contrast media usage should be monitored. Patient injury could occur if the delivery system is not un-flexed prior to removal. Care should be exercised in patients with hypersensitivities to cobalt, nickel, chromium, molybdenum, titanium, manganese, silicon, and/or polymeric materials. The procedure should be conducted under fluoroscopic guidance. Some fluoroscopically guided procedures are associated with a risk of radiation injury to the skin. These injuries may be painful, disfiguring, and long-lasting. Valve recipients should be maintained on anticoagulant/antiplatelet therapy,

Important Safety Information (continued)

except when contraindicated, as determined by their physician. This device has not been tested for use without anticoagulation. Do not add or apply antibiotics to the storage solution (SAPIEN 3 and SAPIEN 3 Ultra only), rinse solution, or to the valve. Balloon valvuloplasty should be avoided in the treatment of failing bioprostheses as this may result in embolization of bioprosthesis material and mechanical disruption of the valve leaflets. Do not perform stand-alone balloon aortic valvuloplasty procedures in the INSPIRIS RESILIA aortic valve for the sizes 19-25 mm. This may expand the valve causing aortic incompetence, coronary embolism or annular rupture. Transcatheter valve replacement in mitral annuloplasty rings is not recommended in cases of partial annuloplasty ring dehiscence due to high risk of PVL. Transcatheter valve replacement in mitral annuloplasty rings is not recommended in cases of partial (incomplete) annuloplasty rings in the absence of annular calcium due to increased risk of valve embolization. Transcatheter valve replacement in mitral annuloplasty rings is not recommended in cases of rigid annuloplasty rings due to increased risk of PVL or THV deformation.

Precautions: Long-term durability has not been established for the valve. Regular medical follow-up is advised to evaluate valve performance. Limited clinical data are available for transcatheter aortic valve replacement in patients with a congenital bicuspid aortic valve who are deemed to be at low surgical risk. Anatomical characteristics should be considered when using the valve in this population. In addition, patient age should be considered as long-term durability of the valve has not been established. Data on TAVR in patients with asymptomatic severe aortic stenosis are based on study of predominantly low surgical risk patients. Limited clinical data to inform benefit-risk considerations are available for TAVR in patients with asymptomatic severe aortic stenosis who are deemed to be at intermediate or greater surgical risk. Glutaraldehyde may cause irritation of the skin, eyes, nose, and throat. Avoid prolonged or repeated exposure to, or breathing of, the solution. Use only with adequate ventilation. If skin contact occurs, immediately flush the affected area with water; in the event of contact with eyes, seek immediate medical attention. For more information about glutaraldehyde exposure, refer to the Safety Data Sheet available from Edwards Lifesciences. If a significant increase in resistance occurs when advancing the catheter through the vasculature, stop advancement and investigate the cause of resistance before proceeding. Do not force passage, as this could increase the risk of vascular complications. As compared to SAPIEN 3, system advancement force may be higher with the use of SAPIEN 3 Ultra/SAPIEN 3 Ultra RESILIA THV in tortuous/challenging vessel anatomies. To maintain proper valve leaflet coaptation, do not overinflate the deployment balloon. Appropriate antibiotic prophylaxis is recommended post-procedure in patients at risk for prosthetic valve infection and endocarditis. Additional precautions for transseptal replacement of a failed mitral valve bioprosthesis include, the presence of devices or thrombus or other abnormalities in the caval vein precluding safe transvenous femoral access for transseptal approach; and the presence of an Atrial Septal Occluder Device or calcium preventing safe transseptal access. Special care must be exercised in mitral valve replacement to avoid entrapment of the subvalvular apparatus. Safety and effectiveness have not been established for patients with the following characteristics/comorbidities: non-calcified aortic annulus; severe ventricular dysfunction with ejection fraction < 20%; congenital unicuspid aortic valve; pre-existing prosthetic ring in the tricuspid position; severe mitral annular calcification (MAC); severe (> 3+) mitral insufficiency, or Gorlin syndrome; blood dyscrasias defined as leukopenia (WBC < 3000 cells/mL), acute anemia (Hb < 9 g/dL), thrombocytopenia (platelet count < 50,000 cells/mL), or history of bleeding diathesis or coagulopathy; hypertrophic cardiomyopathy with or without obstruction (HOCM); echocardiographic evidence of intracardiac mass, thrombus, or vegetation; a known hypersensitivity or contraindication to aspirin, heparin, ticlopidine (Ticlid), or clopidogrel (Plavix), or sensitivity to contrast media, which cannot be adequately premedicated; significant aortic disease, including abdominal aortic or thoracic aneurysm defined as maximal luminal diameter 5 cm or greater, marked tortuosity (hyperacute bend), aortic arch atheroma (especially if thick [> 5 mm], protruding, or ulcerated) or narrowing (especially with calcification and surface irregularities) of the abdominal or thoracic aorta, severe “unfolding” and tortuosity of the thoracic aorta; access characteristics that would preclude safe placement of the Edwards sheath, such as severe obstructive calcification or severe tortuosity; bulky calcified aortic valve leaflets in close proximity to coronary ostia; a concomitant paravalvular leak where the failing prosthesis is not securely fixed in the native annulus or is not structurally intact (e.g., wireform frame fracture, annuloplasty ring dehiscence); or a partially detached leaflet of the failing bioprosthesis that in the aortic position may obstruct a coronary ostium. For Left axillary approach, a left subclavian takeoff angle $\sim \geq 90^\circ$ from the aortic arch causes sharp angles, which may be responsible for potential sheath kinking, subclavian/axillary dissection and aortic arch damage. For left/right axillary approach, ensure there is flow in Left Internal Mammary Artery (LIMA)/Right Internal Mammary Artery (RIMA) during procedure and monitor pressure in homolateral radial artery. Residual mean gradient may be higher in a “THV-in-failing prosthesis” configuration than that observed following implantation of the valve inside a native aortic annulus using the same size device. Patients with elevated mean gradient post procedure should be carefully followed. It is important that the manufacturer, model and size of the preexisting prosthesis be determined, so that the appropriate valve can be

Important Safety Information (continued)

implanted and a prosthesis-patient mismatch be avoided. Additionally, pre-procedure imaging modalities must be employed to make as accurate a determination of the inner diameter as possible.

Potential Adverse Events: Potential risks associated with the overall procedure, including potential access complications associated with standard cardiac catheterization, balloon valvuloplasty, the potential risks of conscious sedation and/or general anesthesia, and the use of angiography: death; stroke/transient ischemic attack, clusters, or neurological deficit; paralysis; permanent disability; respiratory insufficiency or respiratory failure; hemorrhage requiring transfusion or intervention; cardiovascular injury including perforation or dissection of vessels, ventricle, atrium, septum, myocardium, or valvular structures that may require intervention; pericardial effusion or cardiac tamponade; thoracic bleeding; embolization including air, calcific valve material, or thrombus; infection including septicemia and endocarditis; heart failure; myocardial infarction; renal insufficiency or renal failure; conduction system defect which may require a permanent pacemaker; arrhythmia; retroperitoneal bleed; arteriovenous (AV) fistula or pseudoaneurysm; reoperation; ischemia or nerve injury or brachial plexus injury; restenosis; pulmonary edema; pleural effusion; bleeding; anemia; abnormal lab values (including electrolyte imbalance); hypertension or hypotension; allergic reaction to anesthesia, contrast media, or device materials; hematoma; syncope; pain or changes (e.g., wound infection, hematoma, and other wound care complications) at the access site; exercise intolerance or weakness; inflammation; angina; heart murmur; and fever. Additional potential risks associated with the use of the valve, delivery system, and/or accessories include: cardiac arrest; cardiogenic shock; emergency cardiac surgery; cardiac failure or low cardiac output; coronary flow obstruction/transvalvular flow disturbance; device thrombosis requiring intervention; valve thrombosis; device embolization; device migration or malposition requiring intervention; left ventricular outflow tract obstruction; valve deployment in unintended location; valve stenosis; structural valve deterioration (wear, fracture, calcification, leaflet tear/tearing from the stent posts, leaflet retraction, suture line disruption of components of a prosthetic valve, thickening, stenosis); device degeneration; paravalvular or transvalvular leak; valve regurgitation; hemolysis; device explants; nonstructural dysfunction; mechanical failure of delivery system and/or accessories; and non-emergent reoperation.

Thank you



Important Safety Information (continued)

Edwards Crimper

Indications: The Edwards crimper is indicated for use in preparing the Edwards SAPIEN 3 transcatheter heart valve, Edwards SAPIEN 3 Ultra transcatheter heart valve, and the Edwards SAPIEN 3 Ultra RESILIA transcatheter heart valve for implantation.

Contraindications: There are no known contraindications.

Warnings: The device is designed, intended, and distributed for single use only. **Do not resterilize or reuse the device.** There are no data to support the sterility, nonpyrogenicity, and functionality of the device after reprocessing. Do not mishandle the device. Do not use the device if the packaging or any components are not sterile, have been opened or are damaged, or the expiration date has elapsed.

Precautions: For special considerations associated with the use of the Edwards crimper prior to THV implantation, refer to the THV Instructions for Use.

Potential Adverse Events: There are no known potential adverse events associated with the Edwards crimper.

CAUTION: Federal (United States) law restricts these devices to sale by or on the order of a physician.

Edwards, Edwards Lifesciences, the stylized E logo, Edwards SAPIEN, Edwards SAPIEN 3, Edwards SAPIEN 3 Ultra, INSPIRIS, INSPIRIS RESILIA, PARTNER, PARTNER II, PARTNER 3, RESILIA, SAPIEN, SAPIEN 3, and SAPIEN 3 Ultra are trademarks or service marks of Edwards Lifesciences Corporation or its affiliates. All other trademarks are the property of their respective owners.

© 2025 Edwards Lifesciences Corporation. All rights reserved. PP--US-11811 v2.0

Edwards Lifesciences • One Edwards Way, Irvine CA 92614 USA • edwards.com



Edwards